

Positioning Checklist for the Rifton Size 1 Stander

Note: Not all product components are included. Do not use this checklist for product component selection.

Use this Positioning Checklist as a convenient way to ensure optimal use of the Rifton Size 1 Stander. Write notes to customize your instructions for each individual.

INDIVIDUAL'S NAME _____

STAFF NAME _____

DATE _____

THIS POSITIONING CHECKLIST IS INTENDED FOR CONSISTENT USE OF ADAPTIVE EQUIPMENT FOLLOWING ASSESSMENT BY A HEALTH PROFESSIONAL. THE CHECKLIST DOES NOT REPLACE AN INITIAL CLINICAL EVALUATION FOR CORRECT SIZING, COMPONENT SELECTION, AND APPROPRIATE THERAPEUTIC POSITIONING FOR THE INDIVIDUAL. SOME COMPONENT SIZING IS DEPENDENT ON THE MODEL SIZE.

*Positioning references client's right and left sides.

Configuration

- Prone
- Supine



Prone



Supine

Thigh length

Use indexed numbers on Stander.



Right _____ Left _____

Lower leg length

Use indexed numbers on Stander.



Right _____ Left _____

Abduction position

Use indexed degrees on Stander.



Right _____ Left _____

Sandal positioning



Right: Comments _____

Left: Comments _____

Trunk laterals



Upper trunk laterals

Right

Left

Comments _____


Lower trunk laterals

Right

Left

Comments _____

Thigh supports



Medial thigh supports

Right

Left

Comments _____

Lateral thigh supports

Right

Left

Comments _____

Tray



Comments _____

PRONE CONFIGURATION – SIZE 1

Critical supports in prone position



- Trunk strap
 - Pelvic harness
 - Sandals
-

Trunk strap options – Required/critical strap

Standard trunk strap



- Upper attachment point on trunk board
- Center attachment point on trunk board
- Loose
- Snug

OR

Rotation control trunk strap

*Main body strap should be snugly velcroed around client's torso.



- Upper attachment point on trunk board
 - Center attachment point on trunk board
 - Rotation control on Right (snug right strap attachment more than left strap)
 - Rotation control on Left (snug left strap attachment more than right strap)
-

Pelvic harness – Required/critical strap

*For correct use and ideal support, tighten the lower side straps first to get the harness low and snug under the user's hips before adjusting the other straps.

*Secure fifth strap for safety of client.



Prone knee cuffs



- Right
- Left

Comments _____

Stander inclination/tilt



Use indexed degrees on Stander.

SUPINE CONFIGURATION – SIZE 1

Critical supports in supine position



- Trunk strap
- Hip strap
- Knee supports
- Sandals

Trunk support height



Measure from hip tag to top of head support.

Trunk strap options – Required/critical strap



Standard trunk strap

- Upper attachment point on trunk board
- Center attachment point on trunk board
- Loose
- Snug

OR

Rotation control trunk strap

*Main body strap should be snugly velcroed around client's torso.



- Upper attachment point on trunk board
- Center attachment point on trunk board
- Rotation control on Right (snug right strap attachment more than left strap)
- Rotation control on Left (snug left strap attachment more than right strap)

OR

Butterfly harness

*Secure fifth strap for safety of client.



Comments _____

Hip strap options – Required/critical strap



Standard hip strap

- Center attachment point on trunk board
- Lower attachment point on trunk board
- Loose
- Snug

OR

Rotation control hip strap

*Main body strap should be snugly velcroed around client's torso.



- Center attachment point on trunk board
- Lower attachment point on trunk board
- Rotation control on Right (snug right strap attachment more than left strap)
- Rotation control on Left (snug left strap attachment more than right strap)

Supine knee straps – Required/critical strap



Comments _____

Head laterals



- Right
- Left

Width _____

Comments _____

Stander inclination/tilt



Use indexed degrees on Stander.
